

TECHNOVATION ANALYTICAL INSTRUMENTS PVT. LTD.

W-312, 13th Cross Road, M.I.D.C., Rabale, Navi Mumbai-400701, Maharashtra, India. Tel.: 9967177429 / +91-22-27691238

E-mail: taiplem@gmail.com Website: www.technovationindia.com



LOOP POWERED TWO WIRE INDUSTRY STANDARD TRANSMITTERS 4 -20mA

THE INSTRUMENT

Often it is more cost effective to replace a sensing head with one which is pre-calibrated at a testing station, thus reducing disruption of the measuring system. Technovation offers such an industry standard 4-20mA transmitter consisting of an electrochemical sensor, its associated circuitry and in a weather-proof housing.

FEATURES

The two-wire transmitter allows the signal and power to be supplied on a single wire-pair by modulating the power supply current with the input signal source. The transmitter is immune to voltage drops from long runs and noise from motors, relays, actuators, switches, transformers and industrial equipment.

APPLICATIONS

Every transmitter is supplied pre-calibrated for a customer specific range. The table below gives an indication of the gases/ranges offered. These transmitters can be connected to a DC system which supplies the DC Power for the loop. They could also be connected to stand alone controllers and PLCs and other data acquisition systems

STANDARD SPECIFICATIONS

Power Supply : 12 to 28 VDC Single

ended

Sensor Expected life : 2 years in air Sensor type :Electrochemical

Duty : Continuous Operating temp : 20 °C to 45 °C Operating Humidity : 10 - 90 % RH

Operating Pressure · Ambient : 2 M Ohms Output impedance

Calibration : Via built in Span and Zero potentiometers.

: Poly Carbonate

Housing weather proof housing

: W 12 X H 8 X D 8.5

Enclosure Dim

Cms

Sampling : Diffusion mode OR in

Flow through mode

MODULAR CONVENIENT AND INDUSTRY STANDARD COMPATIBLE



TECHNOVATION® 4 - 20 mA TWO WIRE TRANSMITTER

POWER SUPPLY RESTRICTIONS

As the minimum operating voltage for the transmitter is 12 VDC it will impose constraints on the total loop resistance in the external circuit, and this must be taken into account when choosing the supply voltage. This includes the cable resistance, and the measuring resistor at the remote receiver and any meters used for calibration

TABLE OF GASES & RANGES	
OXYGEN	0 - 25 % OR 0-100%
CARBON-MONOXIDE	0 - 1000 PPM
HYDROGEN SULPHIDE	0 - 100 PPM
SULPHUR DIOXIDE	0 - 20 PPM
CHLORINE	0 - 50 PPM
NITROGEN DIOXIDE	0 - 100 PPM
HYDROGEN	0 - 1000 PPM or 2 %
ARSINE	0 - 1PPM
HYDROGEN CYANIDE	0 - 30 PPM
OZONE	0- 1PPM
PHOSPHINE	0 - 1 PPM
AMMONIA	0 - 100 PPM
PHOSGENE	0 - 1PPM
HYDRAZINE	0 - 1PPM
OTHER GASES AND RANGES ON REQUEST	